

**SETSCO SERVICES PTE LTD**

18 Teban Gardens Crescent  
Singapore 608925  
Tel : (65) 6566 7777  
Fax: (65) 6566 7718  
Website: www.setsco.com  
Business Reg. No. 196900269D

## TEST REPORT

(This Report is issued subject to the terms & conditions set out below)

Your Ref: -

Our Ref: MM-25204/WC/3

date: 29<sup>th</sup> October 2011

Page 1 of 9

**Subject** : Witnessing of regulation hydraulic characteristics test on control valves submitted by Lingjack Engineering Works Pte Ltd.

**Tested for** : LINGJACK ENGINEERING WORKS PTE LTD  
No.1 Woodlands Terrace #03-01  
Lingjack industrial building  
Singapore 738471

**Attn:** Mr. Kenneth Lim

**Date & Place of Test** : 28<sup>th</sup> October 2011 at Lingjack Engineering Works Pte Ltd

**Method of Test** : Based on BS EN 1074-5:2000

- 1) Control valves providing flow regulation function (5.3.2.1)
- 2) Control valves providing pressure regulation function (5.3.2.2)
- 3) Mechanical Strength
  - a. Resistance to internal pressure of the shell and of all pressure containing components (5.1.1)
  - b. Resistance to differential pressure (5.1.2)
- 4) Leak-tightness to internal pressure (5.2.1.1)

**Description of Sample** : Two (02) pillars fire hydrant were tested as follows:

| S/No. | Sample Reference | Model    | Size | Qty |
|-------|------------------|----------|------|-----|
| 1     | Control Valve    | C 25-PRV | 1"   | 01  |
| 2     |                  | DF1      | 2"   | 01  |

**Terms & conditions:**

- (1) The Report is prepared for the sole use of the Client and is prepared based upon the Item submitted, the services required by the Client and the conditions under which the Services are performed by SETSCO. The Report is not intended to be representative of similar or equivalent Services on similar or equivalent Items. The Report does not constitute an endorsement by SETSCO of the item.
- (2) SETSCO agrees to use reasonable diligence in the performance of the Services but no warranties are given and none may be implied directly or indirectly relating to the Services, the Report or the facilities of SETSCO.
- (3) The Report may not be used in any publicity material without the written consent of SETSCO.
- (4) The Report may not be reproduced in part or in full unless approval in writing has been given by SETSCO.
- (5) SETSCO shall under no circumstances be liable to the Client or its agents, servants or representatives, in contract, tort (including negligence or breach of statutory duty) or otherwise for any direct or indirect loss or damage suffered by the Client, its agents, servants or representative howsoever arising or whether connected with the Services provided by SETSCO herein.

**Results :**

- 1) Control Valves Providing Flow Regulation Function  
Refer to Table 1
- 2) Control Valves Providing Pressure Regulation Function  
Refer to Table 2
- 3) Mechanical Strength  
(Resistance to internal pressure of the shell and of all pressure containing components)  
Refer to Table 3
- 4) Mechanical Strength Test  
(Resistance to Differential Pressure)  
Refer to Table 4
- 5) Leak-tightness to internal pressure  
Refer to Table 5



Wang Chao  
Testing Officer




MOHAMMAD AZAM  
Engineer (Mechanical Testing)



CHEN YU  
Senior Engineer (Mechanical Testing)  
Mechanical Technology Division

**Results:****Table 1: Flow regulation function**

| Sample Reference | Size | Inlet pressure (bar) | Out let pressure (bar) | Pressure Drop (bar) | Flow rate (l/min) |
|------------------|------|----------------------|------------------------|---------------------|-------------------|
| Control Valve    | 1"   | 6                    | 4                      | 2                   | 71                |
|                  |      | 8                    | 4                      | 4                   | 123               |
|                  |      | 10                   | 4                      | 6                   | 142               |
|                  |      | 12                   | 4                      | 8                   | 172               |
|                  | 2"   | 6                    | 4                      | 2                   | 318               |
|                  |      | 8                    | 4                      | 4                   | 486               |
|                  |      | 10                   | 4                      | 6                   | 604               |
|                  |      | 12                   | 4                      | 8                   | 704               |

**Note:**


- 1) Pressure Gauge (Brand: WiKa, 0~20 bar) Certificate number: NP 1052/11 (Non Singlas Report) dated 10 May 2011
- 2) Pressure Gauge (Brand: WiKa, 0~60 bar) Certificate number: NP 0969/11 (Non Singlas Report) dated 14 Apr 2011
- 3) The wind speed of the environment is not measured.





**Results:****Table 1: Pressure regulation function**

| Sample Reference | Size | Inlet pressure (bar) | Out let pressure (bar) | Flow rate (l/min) |
|------------------|------|----------------------|------------------------|-------------------|
| Control Valve    | 1"   | 13                   | 9                      | 0                 |
|                  |      | 13                   | 4.5                    | 76                |
|                  |      | 13                   | 4                      | 128               |
|                  |      | 13                   | 3                      | 177               |
|                  |      | 13                   | 2                      | 202               |
|                  | 2"   | 13                   | 9                      | 0                 |
|                  |      | 13                   | 6                      | 215               |
|                  |      | 13                   | 4                      | 340               |
|                  |      | 13                   | 2                      | 531               |

**Note:**

- 1) Pressure Gauge (Brand: WiKa, 0~20 bar) Certificate number: NP 1052/11 (Non Singlas Report) dated 10 May 2011
- 2) Pressure Gauge (Brand: WiKa, 0~60 bar) Certificate number: NP 0969/11 (Non Singlas Report) dated 14 Apr 2011
- 3) The wind speed of the environment is not measured.





**Results:**

**Table 3: Mechanical Strength Test**  
**(Resistance to internal pressure of the shell and of all pressure containing components)**

| Sample Reference             | Test Duration (min) | Test Pressure (bar) | Results                       | BS EN 1074-5: 2000 Requirements           |
|------------------------------|---------------------|---------------------|-------------------------------|-------------------------------------------|
| 1" Control Valve (PRV PN 16) | 10                  | 25                  | No leakage & defects observed | No visible leakage & defects is permitted |
| 2" Control Valve (PRV PN 16) |                     |                     | No leakage & defects observed |                                           |

**Table 4: Mechanical Strength Test**  
**(Resistance to Differential Pressure)**

| Sample Reference             |                              | Inlet Test Pressure (bar) | Results             | BS EN 1074-5: 2000 Requirements    |
|------------------------------|------------------------------|---------------------------|---------------------|------------------------------------|
| 1" Control Valve (PRV PN 16) | Obturator in Closed Position | 35                        | No defects observed | Shall show no damage on the valves |
| 2" Control Valve (PRV PN 16) |                              |                           | No defects observed |                                    |

**Table 5: Leak-tightness to internal pressure**

| Sample Reference             | Test Duration (min) | Test Pressure (bar) | Results    | BS EN 1074-5: 2000 Requirements |
|------------------------------|---------------------|---------------------|------------|---------------------------------|
| 1" Control Valve (PRV PN 16) | 10                  | 25                  | No leakage | Leak-tight                      |
| 2" Control Valve (PRV PN 16) |                     |                     | No leakage |                                 |

Note: Pressure Gauge (MIGIAHITA - 0 to 300 kg/cm<sup>2</sup>)

Certificate No. NP 0972/11


The calibration report attached in Appendix is a non-Singlas report.

The torque required to close the obturator was not measured.






Appendix



Photographs shows 2" Control Valve



Photographs shows 1" Control Valve







## Appendix

**NORTHLAB PTE LTD**

No. 27 Woodlands Ind. Park E1, #02-08, Hiangle Ind. Building,  
Singapore 757718. Tel: (65) 6367 0800 Fax: (65) 6367 9367  
Email: northlab.sg@northlab.biz Co. Regn. No. 1996061272  
An ISO 9001, API Spec Q1, ISO TS 29001 & US Navy Certified Organisation

*Certificate of Calibration*

|                      |                                                                                                                                         |                             |                      |
|----------------------|-----------------------------------------------------------------------------------------------------------------------------------------|-----------------------------|----------------------|
| <b>Cert. No.</b>     | <b>: NP 1052/11</b>                                                                                                                     | <b>Job No.</b>              | <b>: C 0947/11</b>   |
| <b>Date of Issue</b> | <b>: 10-May-11</b>                                                                                                                      | <b>Page</b>                 | <b>: 1 of 1</b>      |
| <b>Customer</b>      | <b>: LINGJACK ENGINEERING WORKS<br/>PTE LTD<br/>NO 1 Woodlands Terrace<br/>#03-01 Lingjack Industrial Building<br/>Singapore 738471</b> | <b>Date Received</b>        | <b>: 04-May-11</b>   |
|                      |                                                                                                                                         | <b>Date Calibrated</b>      | <b>: 06-May-11</b>   |
|                      |                                                                                                                                         | <b>Recommended Due Date</b> | <b>: 06-May-12</b>   |
|                      |                                                                                                                                         | <b>Ambient Temperature</b>  | <b>: 20±2°C</b>      |
|                      |                                                                                                                                         | <b>Relative Humidity</b>    | <b>: 55±10% r.h.</b> |
| <b>Instrument</b>    | <b>: <u>PRESSURE GAUGE</u></b>                                                                                                          | <b>Serial No.</b>           | <b>: ---</b>         |
|                      |                                                                                                                                         | <b>(Tag / ID No.)</b>       | <b>: ---</b>         |
| <b>Manufacturer</b>  | <b>: WIKA</b>                                                                                                                           | <b>Range</b>                | <b>: 0 ~ 20 bar</b>  |
| <b>Model No.</b>     | <b>: ---</b>                                                                                                                            |                             |                      |

NorthLab's organisation and practices are derived from ISO/IEC 17025. It therefore meets the relevant requirements of ISO 9001 when acting as the supplier providing test / calibration results. The described instrument has been checked and calibrated at NorthLab under the ambient conditions stated above.

**METHOD OF CALIBRATION**

- The method of calibration is as per method no. P17 in NLQM/001/10 of the Quality Manual.

|                                 |            |                     |                 |
|---------------------------------|------------|---------------------|-----------------|
| <b>Standard instrument used</b> | <b>S/N</b> | <b>Traceability</b> | <b>Due Date</b> |
| 1 Pneumatic Pressure Calibrator | 110        | NMC Singapore       | 15-Oct-11       |

**RESULTS OF CALIBRATION**

| APPLIED<br>EQUIVALENT<br>VALUE<br>(bar) | INSTRUMENT READING (bar) |       |            |       |             |       |            |       |
|-----------------------------------------|--------------------------|-------|------------|-------|-------------|-------|------------|-------|
|                                         | INCREASING               |       |            |       | DECREASING  |       |            |       |
|                                         | Before Adj.              |       | After Adj. |       | Before Adj. |       | After Adj. |       |
|                                         | Rdg                      | Error | Rdg        | Error | Rdg         | Error | Rdg        | Error |
| 0.000                                   | 0.00                     | 0.000 | ---        | ---   | 0.00        | 0.000 | ---        | ---   |
| 5.000                                   | 5.00                     | 0.000 | ---        | ---   | 5.00        | 0.000 | ---        | ---   |
| 10.000                                  | 10.10                    | 0.100 | ---        | ---   | 10.10       | 0.100 | ---        | ---   |
| 15.000                                  | 15.00                    | 0.000 | ---        | ---   | 15.00       | 0.000 | ---        | ---   |
| 20.000                                  | 20.00                    | 0.000 | ---        | ---   | 20.00       | 0.000 | ---        | ---   |

- Accuracy of the instrument is within  $\pm 1\%$  of full scale.
- The expanded uncertainty of measurement is  $\pm 0.029$  bar estimated at a level of confidence of approximately 95% with a coverage factor  $k = 2$ .

**REMARKS / DEVIATIONS IF ANY**

- The recalibration interval should be determined based on the user's requirements.
- The user should determine the suitability of the instrument for its intended use.

  
N. Palaniappan  
Calibration Officer

  
M. Muthiah  
Approved Signatory

This report shall not be reproduced except in full, without the prior written approval of the head of NORTHLAB PTE LTD.

**Pressure Gauge Calibration Certificate**




## Appendix

**NORTHLAB PTE LTD**

No. 27 Woodlands Ind. Park U1, #02-08, Hlaogkie Ind. Building,  
Singapore 757718. Tel: (65) 6367 0800 Fax: (65) 6367 9367  
Email: northlab.sg@northlab.biz Co. Regn. No: 1996061272  
An ISO 9001, API Spec Q1, ISO TS 29001 & US Navy Certified Organisation

*Certificate of Calibration*

|                      |                                                                                                                                         |                             |                       |
|----------------------|-----------------------------------------------------------------------------------------------------------------------------------------|-----------------------------|-----------------------|
| <b>Cert. No.</b>     | <b>: NP 0969/11</b>                                                                                                                     | <b>Job No.</b>              | <b>: C 0766/11</b>    |
| <b>Date of Issue</b> | <b>: 14-Apr-11</b>                                                                                                                      | <b>Page</b>                 | <b>: 1 of 1</b>       |
| <b>Customer</b>      | <b>: LINGJACK ENGINEERING WORKS<br/>PTE LTD<br/>NO 1 Woodlands Terrace<br/>#03-01 Lingjack Industrial Building<br/>Singapore 738471</b> | <b>Date Received</b>        | <b>: 07-Apr-11</b>    |
|                      |                                                                                                                                         | <b>Date Calibrated</b>      | <b>: 13-Apr-11</b>    |
|                      |                                                                                                                                         | <b>Recommended Due Date</b> | <b>: 13-Apr-12</b>    |
|                      |                                                                                                                                         | <b>Ambient Temperature</b>  | <b>: 20±2°C</b>       |
|                      |                                                                                                                                         | <b>Relative Humidity</b>    | <b>: 55±10% r.h.</b>  |
| <b>Instrument</b>    | <b>: PRESSURE GAUGE</b>                                                                                                                 | <b>Serial No.</b>           | <b>: 202654-9-26</b>  |
|                      |                                                                                                                                         | <b>(Tag / ID No. )</b>      | <b>: --- P6/08111</b> |
| <b>Manufacturer</b>  | <b>: WIKA</b>                                                                                                                           | <b>Range</b>                | <b>: 0 ~ 60 bar</b>   |
| <b>Model No.</b>     | <b>: ---</b>                                                                                                                            |                             |                       |

NorthLab's organisation and practices are derived from ISO/IEC 17025. It therefore meets the relevant requirements of ISO 9001 when acting as the supplier providing test / calibration results. The described instrument has been checked and calibrated at NorthLab under the ambient conditions stated above.

**METHOD OF CALIBRATION**

- The method of calibration is as per method no. P03 in NLQM/001/10 of the Quality Manual.

|                                  |            |                     |                 |
|----------------------------------|------------|---------------------|-----------------|
| <b>Standard instrument used</b>  | <b>S/N</b> | <b>Traceability</b> | <b>Due Date</b> |
| 1 Oil Deadweight Pressure Tester | 26412      | NMC Singapore       | 29-May-14       |

**RESULTS OF CALIBRATION**

| APPLIED<br>EQUIVALENT<br>VALUE<br>(bar) | INSTRUMENT READING ( bar ) |       |            |       |             |       |            |       |
|-----------------------------------------|----------------------------|-------|------------|-------|-------------|-------|------------|-------|
|                                         | INCREASING                 |       |            |       | DECREASING  |       |            |       |
|                                         | Before Adj.                |       | After Adj. |       | Before Adj. |       | After Adj. |       |
|                                         | Rdg                        | Error | Rdg        | Error | Rdg         | Error | Rdg        | Error |
| 0.00                                    | 0.0                        | 0.00  | ---        | ---   | 0.0         | 0.00  | ---        | ---   |
| 9.97                                    | 10.0                       | 0.03  | ---        | ---   | 10.2        | 0.23  | ---        | ---   |
| 19.95                                   | 20.0                       | 0.05  | ---        | ---   | 20.2        | 0.25  | ---        | ---   |
| 29.92                                   | 30.0                       | 0.08  | ---        | ---   | 30.2        | 0.28  | ---        | ---   |
| 39.89                                   | 40.0                       | 0.11  | ---        | ---   | 40.2        | 0.31  | ---        | ---   |
| 59.84                                   | 60.0                       | 0.16  | ---        | ---   | 60.0        | 0.16  | ---        | ---   |

- Accuracy of the instrument is within  $\pm 1\%$  of full scale.
- The expanded uncertainty of measurement is  $\pm 0.061$  bar estimated at a level of confidence of approximately 95% with a coverage factor  $k = 2$ .

**REMARKS / DEVIATIONS IF ANY**

- The recalibration interval should be determined based on the user's requirements.
- The user should determine the suitability of the instrument for its intended use.
- The pressure medium used is distilled water.

  
N. Palaniappan  
Calibration Officer

  
M. Muthiah  
Approved Signatory

This report shall not be reproduced except in full, without the prior written approval of the head of NORTHLAB PTE LTD.

**Pressure Gauge Calibration Certificate**






# VERNON MORRIS

TAKES THE FLOW AND PRESSURE

## DIGITAL FLOW GAUGE CALIBRATION CERTIFICATE

Customer: J. D Fire Limited

Serial Number: Size 2 1/2 "

Reference: VM 20025

Paddlewheel Ref No: S Steel

Digital Display Serial No: 0913045

Certificate of conformity: KEMA No.Ex-98.E.1873 X

Calibration Mode: Litres per minute

Accuracy: ± 1% F.S.D.

Static Assembly: Gun Metal

Range: 0-140mhd Serial Number: 109948

Test Specification: BS EN 837-1 1.6%

Tested against: G23113/1 M71 (Traceable to National Standards.)

We hereby certify that the above Digital Flow Gauge has been tested and conforms to the manufacturers specifications and that the Digital Readout has been calibrated with the correct detail.

Date: 16/04/2009

Signed: A. Rep For Vernon Morris Co.Ltd.



Vernon Morris & Co. Ltd.

Chester Road . Bretton . Chester CH4 0DH

Tel: (01244) 660794 2 lines

SALES 0 SERVICE 0 SPARES

Fax: (01244) 681291

Flow Meter Calibration Certificate

